

NIPRO MEDICAL (INDIA) PVT. LTD
 305,215 Atrium Near Hotel Courtyard Marriot, Andheri Kurla Road, Chakala, Andheri (E) 400059.
 Tel.: 022 66114545

DATE: 20/03/17

INSTALLATION / SERVICE REPORT

NMI Report No. 222

HOSPITAL DETAILS		MACHINE DETAILS	
Hospital Name : DCDC kidney care Address : (Esic model hospital Barunagar, Albul) Nephrologist : Dr. Arjun Chief Technician : Bio Medical Engg : Dr. Aditi (m) 0886626627		Model : Suddial 55 plus Serial No : Nature of Call : 16 CX 5291 AMC/Warranty/on call : 16 CX 5293 Installation Date : 16 CX 5290 Warranty Expiry Date : 16 CX 5292 Hour Meter Reading : 16 CX 5294 16 CX 5286	
Problem Report Date : Time :		Problem Observed Date : Time : 10 Am	
Err Code Installation & training.		Err Code 16 CX 5289 16 CX 5288 16 CX 5285 16 CX 5287	
Action Taken Mention all 10 Suddial 55 plus Machine are installed. Calibration of all Machine done, checked all Machine in preparation, Rinse mode. All Machine are working ok. AS Lab not available in central electrolyte Report still pending. → training given Basic training to all technician staff.			
Parts used :	Article No.:	Description :	Qty : Price :
Parts to be Replaced	Article No.:	Description :	Qty : Price :
Call Status - Completed		Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Comments (Service Engineer): AS per DCDC management instruction all Machine installed without UPS or stabilized. They arrange stabilized shortly.			
Comments (Unit - in - Charge)			
Signature & Name of Service Engineer Date : 20/3/17 Time : 5:10 pm		Signature of the Nephrologist / Unit - in - Charge Date : 20/3/2017 Time : 5:13 pm	

 CIN - U51507TN2006PTC070002
 E-mail : info@niproindia.com

 Regd. Office Address :
 2nd Floor, GMR Aero Towers,
 Rajiv Gandhi International Airport, Shamshabad,
 Hyderabad: 500409
 Email id : niprocare@niproindia.com