



NIPRO MEDICAL INDIA PVT. LTD.
 2nd Floor, GMR Aero Towers, Rajiv Gandhi International Airport,
 Shamshabad, Hyderabad - 500 108, Telangana
 Phone : +91 40 24599499 / 498

CUSTOMER COPY

25483

Date: 14/04/2020

INSTALLATION / SERVICE REPORT

Report No.:

HOSPITAL DETAILS
 Hospital Name: Civil Hospital Hisar
 Address:
 Nephrologist: MR. Rajesh
 Medical Engg:
 Item reported Date:
 Code:

MACHINE DETAILS
 Model: 557 Serial No.: 160X5408 → 0558
 Nature of Call: Calibration
 AMC / Warranty / on call:
 Installation Date: 20/08/17
 Warranty Expiry Date: 19/08/22
 Hour Meter Reading:
 Problem Observed

(W.Hour)
 09 → 7693
 10 → 8819
 12 → 9037
 160X5345 → 7027
 46 → 8803
 180X10433 → 4570
 190X13039 → 4696

25/09/2018
 24/09/2023

Calibration, PMS

Date: 20/04/19
 19/04/2024

Work Done: Done all m/c
 Temp. → 37.0 C
 Cond. → 14.0 ns/cm
 Now all m/c's working properly due to
 rinse as well as dialyzer mode,

Items Used: Material code: Description: Qty:

Items to be Replaced: Material code: Description: Qty:

Overall Status - Completed Yes No

Comments (Service Engineer):

Comments (Unit-in-Charge):

Signature & Name of Service Engineer
 Signature: [Signature]
 Name: [Name]
 Time:

Seal & Signature of the Nephrologist / Unit-in-charge
 Signature: [Signature]
 Date: 14/12/20 Time:

